

rteres

Area Percent Report

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Sohil
7/20/2020 3:05:16 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071720\
Data File : FE032397.D
Signal(s) : FID1B.CH
Acq On : 17 Jul 2020 14:53
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE070720.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.419	2.381	2.550	BB	1966565	17782504	94.75%	6.203%
2	4.490	4.450	4.541	BB	1865017	18767957	100.00%	6.547%
3	6.505	6.461	6.545	BB	1742574	18614252	99.18%	6.494%
4	8.300	8.250	8.333	BB	1702310	18345629	97.75%	6.400%
5	9.901	9.848	9.937	BB	1551102	18236160	97.17%	6.362%
6	11.343	11.289	11.386	BB	1476504	18296659	97.49%	6.383%
7	12.652	12.600	12.691	BB	1421466	17931582	95.54%	6.255%
8	13.850	13.790	13.892	BB	1346777	17550637	93.51%	6.123%
9	14.746	14.688	14.812	BB	1066497	15030331	80.09%	5.243%
10	14.953	14.894	15.001	BB	1233413	17287470	92.11%	6.031%
11	15.972	15.914	16.016	BB	1192396	16710638	89.04%	5.830%
12	16.922	16.857	16.961	BB	1141069	16458694	87.70%	5.742%
13	17.808	17.747	17.840	BB	1156244	16369769	87.22%	5.711%
14	18.638	18.578	18.675	BB	967725	15466953	82.41%	5.396%
15	19.420	19.356	19.453	BB	886487	12966753	69.09%	4.523%
16	20.156	20.090	20.195	BB	668140	10046521	53.53%	3.505%
17	20.907	20.689	21.061	BV	510557	13303531	70.88%	4.641%
18	21.853	21.766	21.943	BB	304530	7490204	39.91%	2.613%
Sum of corrected areas:						286656244		

FE070720.M Mon Jul 20 12:20:18 2020