

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071723AL\
 Data File : FE043253.D
 Signal(s) : FID1B.ch
 Acq On : 17 Jul 2023 12:27
 Operator : YP\AJ
 Sample : 03613-02DL 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 TR-4-071323-EPH-2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/18/2023
 Supervised By :Ankita Jodhani 07/18/2023

Integration File: autoint1.e
 Quant Time: Jul 18 04:16:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 071323.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 13 03:02:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.901	635004	3.546 ug/mlm
Spiked Amount	50.000	Recovery	= 7.09%
12) S 1-chlorooctadecane (S...	13.342	27800783	211.582 ug/ml
Spiked Amount	50.000	Recovery	= 423.16%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE071723AL\
Data File : FE043253.D
Signal(s) : FID1B.ch
Acq On : 17 Jul 2023 12:27
Operator : YP\AJ
Sample : 03613-02DL 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
FID_E
ClientSampleId :
TR-4-071323-EPH-2DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/18/2023
Supervised By :Ankita Jodhani 07/18/2023

Integration File: autoint1.e
Quant Time: Jul 18 04:16:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 071323.
Quant Title : GC Extractables
QLast Update : Thu Jul 13 03:02:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.901 min
Delta R.T.: -0.010 min
Response: 635004
Conc: 3.55 ug/ml

Instrument :
FID_E
ClientSampleId :
TR-4-071323-EPH-2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/18/2023
Supervised By :Ankita Jodhani 07/18/2023



#12 1-chlorooctadecane (SURR)

R.T.: 13.342 min
Delta R.T.: -0.004 min
Response: 27800783
Conc: 211.58 ug/ml