

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE072020\
 Data File : FE032417.D
 Signal(s) : FID1B.CH
 Acq On : 20 Jul 2020 13:54
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 21 02:10:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE070720.M
 Quant Title :
 QLast Update : Wed Jul 08 12:04:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.745	15938198	52.390 ug/ml
Target Compounds			
2) N-DECANE	4.486	18749745	53.913 ug/ml
3) N-DODECANE	6.503	18888480	54.125 ug/ml
4) N-TETRADECANE	8.298	18775991	54.030 ug/ml
5) N-HEXADECANE	9.899	18798276	53.918 ug/ml
6) N-OCTADECANE	11.341	19006527	53.748 ug/ml
7) N-EICOSANE	12.651	18758314	53.337 ug/ml
8) N-DOCOSANE	13.848	18500185	52.649 ug/ml
10) N-TETRACOSANE	14.952	18320523	51.756 ug/ml
11) N-HEXACOSANE	15.971	17773022	50.648 ug/ml
12) N-OCTACOSANE	16.920	17517565	49.808 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE072020\
Data File : FE032417.D
Signal(s) : FID1B.CH
Acq On : 20 Jul 2020 13:54
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 21 02:10:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE070720.M
Quant Title :
QLast Update : Wed Jul 08 12:04:06 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

