

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2639

MATRIX: Solid

METHOD: 8270E/3541

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5. GC/MS Calibration Requirements.			✓

The Initial Calibration met the Requirements.

The Continuous Calibration met the requirements.

6. Blank Contamination - If yes, list compounds and concentrations in each blank:	✓
7. Surrogate Recoveries Meet Criteria	✓

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

The Surrogate recoveries were met for all analysis except for OU4-TS-38-071725 [2-Fluorobiphenyl - 37%, Terphenyl-d14 - 38%], OU4-TS-38-071725RE [2-Fluorobiphenyl - 37%, Terphenyl-d14 - 37%], Sample reanalyzed to confirm the failure and both run reported. RT2286MS [Terphenyl-d14 - 44%], RT2286MSD [Terphenyl-d14 - 43%], OU4-TS-39-071725 [Terphenyl-d14 - 44%], OU4-TS-40-071725 [Terphenyl-d14 - 52%], OU4-TS-42-071725 [Terphenyl-d14 - 48%], OU4-TS-43-071725 [Terphenyl-d14 - 53%], OU4-TS-44-071725 [Terphenyl-d14 - 41%] and OU4-TS-44-071725 [Terphenyl-d14 - 39%], One base surrogate is allowed to fail as per SOP, therefore no further corrective action was taken.

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**(CONTINUED)**

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The MS {Q2635-01MS} with File ID: BF143180.D recoveries met the requirements for all compounds except for Benzo(b)fluoranthene[133%], due to matrix interference.			
The MSD {Q2635-01MSD} with File ID: BF143181.D recoveries met the requirements for all compounds except for Benzo(b)fluoranthene[133%], Fluoranthene[133%], due to matrix interference.			
The Blank Spike met requirements for all compounds.			
9. Internal Standard Area/Retention Time Shift Meet Criteria		✓	
Comments: The Internal Standards Areas were met for all analysis.			
10. Extraction Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

As per special requirement for this project form-1 are reported in mg/kg.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date