

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE072823AL\
 Data File : FE043595.D
 Signal(s) : FID1B.ch
 Acq On : 28 Jul 2023 13:34
 Operator : YP\AJ
 Sample : 03791-02DL 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_E
ClientSampleId :
 EO-03-072623-EPH-2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2023
 Supervised By :Ankita Jodhani 07/31/2023

Integration File: sample.E
 Quant Time: Jul 28 14:01:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 071323.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 13 03:02:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.883	1078921	6.025 ug/ml
Spiked Amount	50.000	Recovery =	12.05%
12) S 1-chlorooctadecane (S...	13.319	958082	7.292 ug/mlm
Spiked Amount	50.000	Recovery =	14.58%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE072823AL\
 Data File : FE043595.D
 Signal(s) : FID1B.ch
 Acq On : 28 Jul 2023 13:34
 Operator : YP\AJ
 Sample : 03791-02DL 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1


Instrument :
 FID_E
ClientSampleId :
 EO-03-072623-EPH-2DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/31/2023
 Supervised By :Ankita Jodhani 07/31/2023

Integration File: sample.E
 Quant Time: Jul 28 14:01:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 071323.M
 Quant Title : GC Extractables
 QLast Update : Thu Jul 13 03:02:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.883 min
 Delta R.T.: -0.028 min
 Response: 1078921
 Conc: 6.03 ug/ml

Instrument :
 FID_E
 ClientSampleId :
 EO-03-072623-EPH-2DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/31/2023
 Supervised By :Ankita Jodhani 07/31/2023



#12 1-chlorooctadecane (SURR)

R.T.: 13.319 min
 Delta R.T.: -0.027 min
 Response: 958082
 Conc: 7.29 ug/ml m