

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE072920\
 Data File : FE032464.D
 Signal(s) : FID1B.CH
 Acq On : 29 Jul 2020 10:49
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 30 00:48:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE072820.M
 Quant Title :
 QLast Update : Wed Jul 29 02:31:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.731	16163433	50.371 ug/ml
Target Compounds			
2) N-DECANE	4.477	18396180	50.832 ug/ml
3) N-DODECANE	6.492	18470176	50.847 ug/ml
4) N-TETRADECANE	8.287	18410671	50.771 ug/ml
5) N-HEXADECANE	9.887	18494043	50.629 ug/ml
6) N-OCTADECANE	11.329	18785280	50.559 ug/ml
7) N-EICOSANE	12.637	18695402	50.485 ug/ml
8) N-DOCOSANE	13.835	18663185	50.381 ug/ml
10) N-TETRACOSANE	14.938	18760985	50.308 ug/ml
11) N-HEXACOSANE	15.956	18501316	50.180 ug/ml
12) N-OCTACOSANE	16.905	18383338	49.922 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE072920\
Data File : FE032464.D
Signal(s) : FID1B.CH
Acq On : 29 Jul 2020 10:49
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 30 00:48:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE072820.M
Quant Title :
QLast Update : Wed Jul 29 02:31:52 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

