

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE073022\
 Data File : FE036767.D
 Signal(s) : FID1B.ch
 Acq On : 29 Jul 2022 09:21
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/29/2022
 Supervised By :Ankita Jodhani 08/01/2022

Integration File: autoint1.e
 Quant Time: Jul 29 12:03:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE072822.M
 Quant Title :
 QLast Update : Thu Jul 28 04:33:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.934	6181397	50.630 ug/ml
Target Compounds			
1) N-OCTANE	1.952	6829513	50.447 ug/ml
2) N-DECANE	4.484	6979703	52.611 ug/ml
3) N-DODECANE	6.654	7114567	52.364 ug/ml
4) N-TETRADECANE	8.481	7122562	51.764 ug/ml
5) N-HEXADECANE	10.087	7215602	52.026 ug/ml
6) N-OCTADECANE	11.528	7272446	51.608 ug/ml
7) N-EICOSANE	12.838	7277574	51.321 ug/ml
8) N-DOCOSANE	14.036	7096183	50.990 ug/ml
10) N-TETRACOSANE	15.138	7069658	50.484 ug/ml
11) N-HEXACOSANE	16.161	7005010	51.612 ug/ml
12) N-OCTACOSANE	17.111	7001902	52.875 ug/ml
13) N-TRIACONTANE	17.999	7107999	54.774 ug/ml
14) N-DOTRIACONTANE	18.832	6633040	52.271 ug/mlm
15) N-TETRATRIACONTANE	19.614	6425162	52.133 ug/mlm
16) N-HEXATRIACONTANE	20.354	5953219	52.334 ug/mlm
17) N-OCTATRIACONTANE	21.078	5916440	52.287 ug/mlm
18) N-TETRACONTANE	21.964	5888988	47.250 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE073022\
Data File : FE036767.D
Signal(s) : FID1B.ch
Acq On : 29 Jul 2022 09:21
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 07/29/2022
Supervised By : Ankita Jodhani 08/01/2022

Integration File: autoint1.e
Quant Time: Jul 29 12:03:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE072822.M
Quant Title :
QLast Update : Thu Jul 28 04:33:35 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

