

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE073118\
 Data File : FE028448.D
 Signal(s) : FID1B.CH
 Acq On : 01 Aug 2018 13:38
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 01 14:37:50 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE073118.M
 Quant Title :
 QLast Update : Tue Jul 31 14:21:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.993	8311921	51.256 ug/ml
Target Compounds			
2) N-DECANE	4.677	8733018	47.287 ug/ml
3) N-DODECANE	6.705	9440463	47.625 ug/ml
4) N-TETRADECANE	8.512	10078812	48.952 ug/ml
5) N-HEXADECANE	10.122	10390751	50.165 ug/ml
6) N-OCTADECANE	11.572	10365155	50.786 ug/ml
7) N-EICOSANE	12.889	10054998	51.085 ug/ml
8) N-DOCOSANE	14.092	9734804	51.220 ug/ml
10) N-TETRACOSANE	15.201	9432880	50.765 ug/ml
11) N-HEXACOSANE	16.225	9242490	50.537 ug/ml
12) N-OCTACOSANE	17.180	9176931	49.801 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE073118\
Data File : FE028448.D
Signal(s) : FID1B.CH
Acq On : 01 Aug 2018 13:38
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 01 14:37:50 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE073118.M
Quant Title :
QLast Update : Tue Jul 31 14:21:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

