

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080218\
 Data File : FE028468.D
 Signal(s) : FID1B.CH
 Acq On : 02 Aug 2018 23:12
 Operator : AJ\SJ
 Sample : PB111694BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 03 09:31:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE073118.M
 Quant Title :
 QLast Update : Tue Jul 31 14:21:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.978	301197	1.857 ug/ml
Target Compounds				
1)	N-OCTANE	2.567	289536	1.766 ug/ml
2)	N-DECANE	4.667	321583	1.741 ug/ml
3)	N-DODECANE	6.695	346089	1.746 ug/ml
4)	N-TETRADECANE	8.499	367744	1.786 ug/ml
5)	N-HEXADECANE	10.108	364227	1.758 ug/ml
6)	N-OCTADECANE	11.557	357457	1.751 ug/ml
7)	N-EICOSANE	12.872	346496	1.760 ug/ml
8)	N-DOCOSANE	14.076	334373	1.759 ug/ml
10)	N-TETRACOSANE	15.184	362732	1.952 ug/ml
11)	N-HEXACOSANE	16.209	322964	1.766 ug/ml
12)	N-OCTACOSANE	17.163	329316	1.787 ug/ml
13)	N-TRIACONTANE	18.055	342331	1.835 ug/ml
14)	N-DOTRIACONTANE	18.889	341478	1.878 ug/ml
15)	N-TETRATRIACONTANE	19.676	231985	1.341 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080218\
Data File : FE028468.D
Signal(s) : FID1B.CH
Acq On : 02 Aug 2018 23:12
Operator : AJ\SJ
Sample : PB111694BS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 03 09:31:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE073118.M
Quant Title :
QLast Update : Tue Jul 31 14:21:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

