

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080219\
 Data File : FE030216.D
 Signal(s) : FID1B.CH
 Acq On : 02 Aug 2019 15:31
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 02 16:01:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE071719.M
 Quant Title :
 QLast Update : Thu Jul 18 07:59:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.588	8880206	49.737 ug/ml
Target Compounds			
1) N-OCTANE	2.356	67960352	418.013 ug/ml
2) N-NONANE	3.324	70790770	432.177 ug/ml
3) N-DECANE	4.380	108699989	429.209 ug/ml
4) N-UNDECANE	5.409	109269982	639.484 ug/ml
5) N-DODECANE	6.391	108606753	635.887 ug/ml
6) N-TETRADECANE	8.188	109065180	629.853 ug/ml
7) N-HEXADECANE	9.785	90302871	519.301 ug/ml
8) N-OCTADECANE	11.207	44582453	257.552 ug/ml
9) N-EICOSANE	12.498	17629008	102.231 ug/ml
10) N-TETRACOSANE	14.802	17117694	99.855 ug/ml
11) N-OCTACOSANE	16.764	8216508	48.811 ug/ml
12) N-DOTRIACONTANE	18.485	7808529	47.472 ug/ml
14) N-HEXATRIACONTANE	20.011	7321347	48.503 ug/ml
15) N-TETRACONTANE	21.600	7159395	49.728 ug/ml
16) N-TETRATETRAOCTANE	24.202	7177845	46.563 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080219\
Data File : FE030216.D
Signal(s) : FID1B.CH
Acq On : 02 Aug 2019 15:31
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 02 16:01:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE071719.M
Quant Title :
QLast Update : Thu Jul 18 07:59:42 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

