

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080222AL\
 Data File : FE036813.D
 Signal(s) : FID1B.ch
 Acq On : 02 Aug 2022 09:19
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Aug 03 03:35:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 071722.M
 Quant Title : GC Extractables
 QLast Update : Mon Jul 18 02:35:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.834	3333217	17.897 ug/ml
Spiked Amount 50.000		Recovery =	35.79%
12) S 1-chlorooctadecane (S...	13.274	2617244	17.202 ug/ml
Spiked Amount 50.000		Recovery =	34.40%
Target Compounds			
1) T n-Nonane (C9)	3.226	2960273	19.203 ug/ml
2) T n-Decane (C10)	4.480	2965348	19.237 ug/ml
3) T A~Naphthalene (C11.7)	6.180	3257390	19.062 ug/ml
4) T n-Dodecane (C12)	6.648	3016732	19.208 ug/ml
5) T A~2-methylnaphthalene...	7.276	3230683	18.797 ug/ml
6) T n-Tetradecane (C14)	8.474	3029991	18.560 ug/ml
7) T n-Hexadecane (C16)	10.080	3026041	18.359 ug/ml
8) T n-Octadecane (C18)	11.522	3045115	17.918 ug/ml
10) T n-Eicosane (C20)	12.830	2986639	17.558 ug/ml
11) T n-Heneicosane (C21)	13.443	2971197	17.360 ug/ml
13) T n-Docosane (C22)	14.028	2947934	17.266 ug/ml
14) T n-Tetracosane (C24)	15.132	2928026	17.177 ug/ml
15) T n-Hexacosane (C26)	16.152	2790000	16.688 ug/ml
16) T n-Octacosane (C28)	17.103	2710510	16.596 ug/ml
17) T n-Tricontane (C30)	17.991	2541801	16.727 ug/ml
18) T n-Dotriacontane (C32)	18.823	2068195	17.196 ug/ml
19) T n-Tetratriacontane (C34)	19.606	1496867	17.975 ug/ml
20) T n-Hexatriacontane (C36)	20.349	1235594	21.043 ug/ml
21) T n-Octatriacontane (C38)	21.072	1279736	25.524 ug/ml
22) T n-Tetracontane (C40)	21.951	917283	17.762 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080222AL\
Data File : FE036813.D
Signal(s) : FID1B.ch
Acq On : 02 Aug 2022 09:19
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :

Integration File: autoint1.e
Quant Time: Aug 03 03:35:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 071722.M
Quant Title : GC Extractables
QLast Update : Mon Jul 18 02:35:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.226 min
Delta R.T.: -0.013 min
Response: 2960273
Conc: 19.20 ug/ml

Instrument :
FID_E
ClientSampleId :




#2 n-Decane (C10)

R.T.: 4.480 min
Delta R.T.: -0.013 min
Response: 2965348
Conc: 19.24 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 6.180 min
Delta R.T.: -0.013 min
Response: 3257390
Conc: 19.06 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.648 min
Delta R.T.: -0.013 min
Response: 3016732
Conc: 19.21 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.276 min
Delta R.T.: -0.012 min
Response: 3230683
Conc: 18.80 ug/ml

Instrument :
FID_E
ClientSampleId :




#6 n-Tetradecane (C14)

R.T.: 8.474 min
Delta R.T.: -0.013 min
Response: 3029991
Conc: 18.56 ug/ml




#7 n-Hexadecane (C16)

R.T.: 10.080 min
Delta R.T.: -0.014 min
Response: 3026041
Conc: 18.36 ug/ml



#8 n-Octadecane (C18)


R.T.: 11.522 min
Delta R.T.: -0.013 min
Response: 3045115
Conc: 17.92 ug/ml



#9 ortho-Terphenyl (SURR)


R.T.: 11.834 min
Delta R.T.: -0.012 min
Response: 3333217
Conc: 17.90 ug/ml

Instrument :
FID_E
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.830 min
Delta R.T.: -0.013 min
Response: 2986639
Conc: 17.56 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.443 min
Delta R.T.: -0.012 min
Response: 2971197
Conc: 17.36 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 13.274 min
Delta R.T.: -0.012 min
Response: 2617244
Conc: 17.20 ug/ml



#13 n-Docosane (C22)


R.T.: 14.028 min
Delta R.T.: -0.013 min
Response: 2947934
Conc: 17.27 ug/ml

Instrument :
FID_E
ClientSampleId :




#14 n-Tetracosane (C24)

R.T.: 15.132 min
Delta R.T.: -0.012 min
Response: 2928026
Conc: 17.18 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.152 min
Delta R.T.: -0.012 min
Response: 2790000
Conc: 16.69 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.103 min
Delta R.T.: -0.009 min
Response: 2710510
Conc: 16.60 ug/ml



#17 n-Tricontane (C30)


R.T.: 17.991 min
Delta R.T.: -0.008 min
Response: 2541801
Conc: 16.73 ug/ml

Instrument :
FID_E
ClientSampleId :




#18 n-Dotriacontane (C32)

R.T.: 18.823 min
Delta R.T.: -0.009 min
Response: 2068195
Conc: 17.20 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.606 min
Delta R.T.: -0.009 min
Response: 1496867
Conc: 17.97 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.349 min
Delta R.T.: -0.005 min
Response: 1235594
Conc: 21.04 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.072 min
Delta R.T.: -0.006 min
Response: 1279736
Conc: 25.52 ug/ml

Instrument :
FID_E
ClientSampleId :



#22 n-Tetracontane (C40)

R.T.: 21.951 min
Delta R.T.: -0.009 min
Response: 917283
Conc: 17.76 ug/ml