

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080323AL\
 Data File : FE043701.D
 Signal(s) : FID1B.ch
 Acq On : 03 Aug 2023 02:12
 Operator : YP\AJ
 Sample : 03860-06
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 SP-2B

Integration File: autoint1.e
 Quant Time: Aug 03 04:33:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 080323.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 02 09:56:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.944	6396790	31.041 ug/ml
Spiked Amount 50.000		Recovery =	62.08%
12) S 1-chlorooctadecane (S...	13.382	4842747	31.034 ug/ml
Spiked Amount 50.000		Recovery =	62.07%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080323AL\
Data File : FE043701.D
Signal(s) : FID1B.ch
Acq On : 03 Aug 2023 02:12
Operator : YP\AJ
Sample : 03860-06
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
SP-2B

Integration File: autoint1.e
Quant Time: Aug 03 04:33:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 080323.M
Quant Title : GC Extractables
QLast Update : Wed Aug 02 09:56:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.944 min
Delta R.T.: 0.003 min
Response: 6396790
Conc: 31.04 ug/ml

Instrument :
FID_E
ClientSampleId :
SP-2B



#12 1-chlorooctadecane (SURR)

R.T.: 13.382 min
Delta R.T.: 0.003 min
Response: 4842747
Conc: 31.03 ug/ml