

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080423AL\
 Data File : FE043707.D
 Signal(s) : FID1B.ch
 Acq On : 03 Aug 2023 11:18
 Operator : YP\AJ
 Sample : 03860-02
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 SOUTH-WALL

Integration File: autoint1.e
 Quant Time: Aug 04 01:30:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 080323.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 02 09:56:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.943	7115624	34.530 ug/ml
Spiked Amount 50.000		Recovery =	69.06%
12) S 1-chlorooctadecane (S...	13.382	5796624	37.147 ug/ml
Spiked Amount 50.000		Recovery =	74.29%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080423AL\
Data File : FE043707.D
Signal(s) : FID1B.ch
Acq On : 03 Aug 2023 11:18
Operator : YP\AJ
Sample : 03860-02
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
SOUTH-WALL

Integration File: autoint1.e
Quant Time: Aug 04 01:30:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 080323.M
Quant Title : GC Extractables
QLast Update : Wed Aug 02 09:56:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.943 min
Delta R.T.: 0.002 min
Response: 7115624
Conc: 34.53 ug/ml

Instrument :
FID_E
ClientSampleId :
SOUTH-WALL



#12 1-chlorooctadecane (SURR)

R.T.: 13.382 min
Delta R.T.: 0.003 min
Response: 5796624
Conc: 37.15 ug/ml