

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080618\
 Data File : FE028475.D
 Signal(s) : FID1B.CH
 Acq On : 06 Aug 2018 11:35
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 06 13:14:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE073118.M
 Quant Title :
 QLast Update : Tue Jul 31 14:21:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.005	7141860	44.040 ug/ml
Target Compounds			
2) N-DECANE	4.680	8185413	44.321 ug/ml
3) N-DODECANE	6.712	8775034	44.268 ug/ml
4) N-TETRADECANE	8.520	9100074	44.198 ug/ml
5) N-HEXADECANE	10.131	9110474	43.984 ug/ml
6) N-OCTADECANE	11.581	8939493	43.800 ug/ml
7) N-EICOSANE	12.898	8603384	43.710 ug/ml
8) N-DOCOSANE	14.102	8325895	43.807 ug/ml
10) N-TETRACOSANE	15.212	8134912	43.780 ug/ml
11) N-HEXACOSANE	16.238	8091287	44.242 ug/ml
12) N-OCTACOSANE	17.193	8134202	44.142 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE080618\
Data File : FE028475.D
Signal(s) : FID1B.CH
Acq On : 06 Aug 2018 11:35
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 06 13:14:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE073118.M
Quant Title :
QLast Update : Tue Jul 31 14:21:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

