

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD CCC
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE08142
Data File : FE049606.D
Signal(s) : FID1B.ch
Acq On : 14 Aug 2024 20:37
Sample : 20 PPM ALIPHATIC HC STD CCC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/15/2024
Supervised By :Ankita Jodhani 08/15/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 081424.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.277	3.244	3.339	BBA	313431	3064657	91.64%	4.440%
2	4.528	4.458	4.628	BB	325350	3226491	96.48%	4.675%
3	6.234	6.174	6.434	BB	312700	3310693	98.99%	4.797%
4	6.695	6.631	6.778	BB	331295	3244372	97.01%	4.701%
5	7.330	7.288	7.504	PB	285304	3122218	93.36%	4.524%
6	8.521	8.451	8.641	BB	297382	3195093	95.54%	4.629%
7	10.128	10.084	10.251	BB	292980	3247255	97.10%	4.705%
8	11.570	11.531	11.694	BB	294310	3258386	97.43%	4.721%
9	11.887	11.814	11.998	BB	310260	3344356	100.00%	4.845%
10	12.877	12.838	12.968	BB	267531	3092325	92.46%	4.480%
11	13.322	13.248	13.411	BV	207756	2545091	76.10%	3.687%
12	13.489	13.451	13.594	VB	257180	3037249	90.82%	4.400%
13	14.074	14.041	14.161	PB	250705	3008197	89.95%	4.358%
14	15.176	15.098	15.281	BB	232934	3018539	90.26%	4.373%
15	16.196	16.024	16.291	BB	229604	2994772	89.55%	4.339%
16	17.144	17.048	17.244	BB	230815	3032362	90.67%	4.393%
17	18.031	17.951	18.084	PV	239732	3248980	97.15%	4.707%
18	18.862	18.801	18.914	BV	232947	3307559	98.90%	4.792%
19	19.646	19.581	19.738	BV	227369	3292769	98.46%	4.771%
20	20.384	20.341	20.429	BBA	213547	3244489	97.01%	4.701%
21	21.111	21.051	21.169	BBA	180381	3189083	95.36%	4.620%
22	22.007	21.951	22.064	BV	126784	2996208	89.59%	4.341%
Sum of corrected areas:						69021142		

Aliphatic EPH 081424.M Thu Aug 15 03:10:20 2024