

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE081621\
 Data File : FE034154.D
 Signal(s) : FID1B.CH
 Acq On : 16 Aug 2021 11:27
 Operator : DD\AJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 100 TRPH STD

Integration File: autoint1.e
 Quant Time: Aug 16 13:17:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081621.M
 Quant Title :
 QLast Update : Mon Aug 16 13:16:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.676	28776143	94.984 ug/ml
Target Compounds			
1) N-OCTANE	2.379	33585965	94.939 ug/ml
2) N-DECANE	4.428	33885512	95.307 ug/ml
3) N-DODECANE	6.438	34250375	95.348 ug/ml
4) N-TETRADECANE	8.232	34448379	95.293 ug/ml
5) N-HEXADECANE	9.832	34166390	95.413 ug/ml
6) N-OCTADECANE	11.273	34381766	95.575 ug/ml
7) N-EICOSANE	12.583	34803029	95.504 ug/ml
8) N-DOCOSANE	13.781	34532544	95.506 ug/ml
10) N-TETRACOSANE	14.884	33873149	95.645 ug/ml
11) N-HEXACOSANE	15.903	33057446	95.920 ug/ml
12) N-OCTACOSANE	16.851	32109396	96.507 ug/ml
13) N-TRIACONTANE	17.738	30572790	97.389 ug/ml
14) N-DOTRIACONTANE	18.569	28661579	98.279 ug/ml
15) N-TETRATRIACONTANE	19.351	25696520	99.790 ug/ml
16) N-HEXATRIACONTANE	20.089	22943090	99.439 ug/ml
17) N-OCTATRIACONTANE	20.820	20399267	98.920 ug/ml
18) N-TETRACONTANE	21.722	17383188	94.112 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE081621\
Data File : FE034154.D
Signal(s) : FID1B.CH
Acq On : 16 Aug 2021 11:27
Operator : DD\AJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
100 TRPH STD

Integration File: autoint1.e
Quant Time: Aug 16 13:17:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081621.M
Quant Title :
QLast Update : Mon Aug 16 13:16:41 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

