

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE081621\
 Data File : FE034157.D
 Signal(s) : FID1B.CH
 Acq On : 16 Aug 2021 14:22
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 10 TRPH STD

Manual Integrations
 APPROVED

mohammad
 8/17/2021 3:45:22 PM

Integration File: autoint1.e
 Quant Time: Aug 16 14:54:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081621.M
 Quant Title :
 QLast Update : Mon Aug 16 13:16:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.651	2918011	9.923 ug/ml
Target Compounds			
1) N-OCTANE	2.379	3451587	10.058 ug/ml
2) N-DECANE	4.414	3467508	10.085 ug/ml
3) N-DODECANE	6.419	3449174	9.999 ug/ml
4) N-TETRADECANE	8.210	3521098	10.043 ug/ml
5) N-HEXADECANE	9.809	3499712	10.032 ug/ml
6) N-OCTADECANE	11.248	3501799	9.978 ug/ml
7) N-EICOSANE	12.556	3531504	9.944 ug/ml
8) N-DOCOSANE	13.752	3511749	9.946 ug/ml
10) N-TETRACOSANE	14.853	3426959	9.908 ug/ml
11) N-HEXACOSANE	15.874	3327015	9.933 ug/ml
12) N-OCTACOSANE	16.822	3261610	6.498 ug/ml
13) N-TRIACONTANE	17.708	3044580	9.994 ug/ml
14) N-DOTRIACONTANE	18.541	2910837	9.132 ug/ml
15) N-TETRATRIACONTANE	19.323	2496339	10.108 ug/ml
16) N-HEXATRIACONTANE	20.061	2306888	10.562 ug/ml
17) N-OCTATRIACONTANE	20.788	2021377	10.064 ug/ml
18) N-TETRACONTANE	21.683	1732629	9.727 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE081621\
Data File : FE034157.D
Signal(s) : FID1B.CH
Acq On : 16 Aug 2021 14:22
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_E
ClientSampled :
10 TRPH STD

Manual Integrations
APPROVED

mohammad
8/17/2021 3:45:22 PM

Integration File: autoint1.e
Quant Time: Aug 16 14:54:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081621.M
Quant Title :
QLast Update : Mon Aug 16 13:16:41 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

