

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE082118\
 Data File : FE028587.D
 Signal(s) : FID1B.CH
 Acq On : 21 Aug 2018 12:49
 Operator : AJ\SJ
 Sample : 1760 TRPH STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 21 13:18:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE082118.M
 Quant Title :
 QLast Update : Tue Aug 21 13:05:16 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.973	5429059	20.318 ug/mlm
Target Compounds			
1) N-OCTANE	2.564	18084044	151.610 ug/ml
2) N-NONANE	3.593	18923304	149.723 ug/ml
3) N-DECANE	4.674	29554892	148.191 ug/ml
4) N-UNDECANE	5.717	32529493	221.292 ug/ml
5) N-DODECANE	6.704	34082165	222.041 ug/ml
6) N-TETRADECANE	8.515	38134275	226.936 ug/ml
7) N-HEXADECANE	10.126	36496675	194.753 ug/ml
8) N-OCTADECANE	11.564	20284616	99.923 ug/ml
9) N-EICOSANE	12.870	8820667	40.588 ug/ml
10) N-TETRACOSANE	15.184	9180089	40.882 ug/ml
11) N-OCTACOSANE	17.155	4332273	20.537 ug/ml
12) N-DOTRIACONTANE	18.879	3607912	20.882 ug/ml
14) N-HEXATRIACONTANE	20.406	2603715	20.688 ug/ml
15) N-TETRACONTANE	22.230	2272813	20.833 ug/ml
16) N-TETRTETRACOOCTANE	25.369	2111916	21.047 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE082118\
Data File : FE028587.D
Signal(s) : FID1B.CH
Acq On : 21 Aug 2018 12:49
Operator : AJ\SJ
Sample : 1760 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 21 13:18:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE082118.M
Quant Title :
QLast Update : Tue Aug 21 13:05:16 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

