

rteres

Area Percent Report

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
8/25/2020 10:34:28 AM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE082420\
Data File : FE032605.D
Signal(s) : FID1B.CH
Acq On : 24 Aug 2020 13:36
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.385	2.344	2.537	BB	1887329	17838886	97.41%	6.540%
2	4.459	4.417	4.525	BB	1796426	18272340	99.78%	6.699%
3	6.476	6.432	6.527	BB	1730011	18312501	100.00%	6.714%
4	8.272	8.190	8.309	BB	1586299	18223223	99.51%	6.681%
5	9.873	9.821	9.913	BB	1516653	18117497	98.94%	6.643%
6	11.316	11.262	11.358	BB	1444989	18195788	99.36%	6.671%
7	12.625	12.567	12.672	BB	1380528	17851418	97.48%	6.545%
8	13.823	13.767	13.872	BB	1322858	17473671	95.42%	6.407%
9	14.720	14.663	14.775	BB	1022825	14876305	81.24%	5.454%
10	14.927	14.868	14.975	BB	1293878	17164008	93.73%	6.293%
11	15.946	15.887	15.988	BB	1176205	16480807	90.00%	6.043%
12	16.897	16.821	16.932	BB	1099427	16051190	87.65%	5.885%
13	17.783	17.713	17.817	BB	1078367	15679050	85.62%	5.749%
14	18.615	18.550	18.647	BB	969600	14496058	79.16%	5.315%
15	19.394	19.330	19.423	M	803865	11817239	64.53%	4.333%
16	20.132	20.076	20.164	M	575776	8838352	48.26%	3.241%
17	20.880	20.818	20.934	BV	365899	7041990	38.45%	2.582%
18	21.817	21.751	21.866	PV	263295	6014960	32.85%	2.205%
Sum of corrected areas:						272745284		

FE081120.M Tue Aug 25 09:32:53 2020