

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE082820\
 Data File : FE032637.D
 Signal(s) : FID1B.CH
 Acq On : 28 Aug 2020 21:32
 Operator : DD\AJ
 Sample : L3832-03MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 SIMW-02-20200827MSD

Manual Integrations
 APPROVED

Ankita
 8/31/2020 1:18:03 PM

Integration File: Sample.e
 Quant Time: Aug 31 00:56:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
 Quant Title :
 QLast Update : Thu Aug 13 11:46:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.699	6137310	19.070 ug/ml
Target Compounds			
1) N-OCTANE	2.386	4097244	11.755 ug/ml
2) N-DECANE	4.446	5840241	16.265 ug/ml
3) N-DODECANE	6.459	6090884	16.850 ug/ml
4) N-TETRADECANE	8.254	6384290	17.677 ug/ml
5) N-HEXADECANE	9.853	6556269	17.972 ug/ml
6) N-OCTADECANE	11.295	6994245	18.829 ug/ml
7) N-EICOSANE	12.603	7030182	18.979 ug/ml
8) N-DOCOSANE	13.801	6963578	18.774 ug/ml
10) N-TETRACOSANE	14.903	6872628	18.356 ug/ml
11) N-HEXACOSANE	15.922	6621955	17.843 ug/ml
12) N-OCTACOSANE	16.872	6593218	17.699 ug/ml
13) N-TRIACONTANE	17.757	6509481	17.454 ug/ml
14) N-DOTRIACONTANE	18.590	6437220	18.267 ug/ml
15) N-TETRATRIACONTANE	19.373	6329319	21.428 ug/mlm
16) N-HEXATRIACONTANE	20.113	6228820	27.629 ug/ml
17) N-OCTATRIACONTANE	20.858	5910314	32.451 ug/ml
18) N-TETRACONTANE	21.792	5804480	34.829 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE082820\
Data File : FE032637.D
Signal(s) : FID1B.CH
Acq On : 28 Aug 2020 21:32
Operator : DD\AJ
Sample : L3832-03MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
SIMW-02-20200827MSD

Manual Integrations
APPROVED

Ankita
8/31/2020 1:18:03 PM

Integration File: Sample.e
Quant Time: Aug 31 00:56:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
Quant Title :
QLast Update : Thu Aug 13 11:46:52 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

