

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE083023AL\
 Data File : FE044043.D
 Signal(s) : FID1B.ch
 Acq On : 30 Aug 2023 23:04
 Operator : YP\AJ
 Sample : 50 PPM ALIPHATIC HC STD2
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM ALIPHATIC HC STD2

Integration File: autoint1.e
 Quant Time: Aug 31 03:33:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 083023.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 31 03:33:38 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.941	8763172	48.242 ug/ml
Spiked Amount 50.000		Recovery =	96.48%
12) S 1-chlorooctadecane (S...	13.378	6447621	48.294 ug/ml
Spiked Amount 50.000		Recovery =	96.59%
Target Compounds			
1) T n-Nonane (C9)	3.324	7555132	48.556 ug/ml
2) T n-Decane (C10)	4.576	7730632	48.539 ug/ml
3) T A~Naphthalene (C11.7)	6.278	8385029	48.718 ug/ml
4) T n-Dodecane (C12)	6.741	7826228	48.526 ug/ml
5) T A~2-methylnaphthalene...	7.374	8039305	48.588 ug/ml
6) T n-Tetradecane (C14)	8.569	7794750	48.533 ug/ml
7) T n-Hexadecane (C16)	10.178	8007388	48.361 ug/ml
8) T n-Octadecane (C18)	11.624	8043713	48.328 ug/ml
10) T n-Eicosane (C20)	12.933	7643013	48.265 ug/ml
11) T n-Heneicosane (C21)	13.546	7473818	48.230 ug/ml
13) T n-Docosane (C22)	14.132	7387564	48.234 ug/ml
14) T n-Tetracosane (C24)	15.237	7230423	48.191 ug/ml
15) T n-Hexacosane (C26)	16.259	6979817	48.207 ug/ml
16) T n-Octacosane (C28)	17.209	6681836	48.172 ug/ml
17) T n-Tricontane (C30)	18.098	6643215	48.066 ug/ml
18) T n-Dotriacontane (C32)	18.932	6467229	48.032 ug/ml
19) T n-Tetratriacontane (C34)	19.718	6381958	48.175 ug/ml
20) T n-Hexatriacontane (C36)	20.459	6311832	48.339 ug/ml
21) T n-Octatriacontane (C38)	21.204	6425948	48.367 ug/ml
22) T n-Tetracontane (C40)	22.128	6459727	48.711 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE083023AL\
Data File : FE044043.D
Signal(s) : FID1B.ch
Acq On : 30 Aug 2023 23:04
Operator : YP\AJ
Sample : 50 PPM ALIPHATIC HC STD2
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM ALIPHATIC HC STD2

Integration File: autoint1.e
Quant Time: Aug 31 03:33:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 083023.M
Quant Title : GC Extractables
QLast Update : Thu Aug 31 03:33:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

