

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE083120\
 Data File : FE032656.D
 Signal(s) : FID1B.CH
 Acq On : 31 Aug 2020 14:36
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Sep 01 00:37:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
 Quant Title :
 QLast Update : Thu Aug 13 11:46:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.701	15650404	48.630 ug/ml
Target Compounds			
1) N-OCTANE	2.379	16765844	48.100 ug/ml
2) N-DECANE	4.447	19072862	53.118 ug/ml
3) N-DODECANE	6.463	19144323	52.962 ug/ml
4) N-TETRADECANE	8.257	18994645	52.593 ug/ml
5) N-HEXADECANE	9.857	18893634	51.792 ug/ml
6) N-OCTADECANE	11.298	18983504	51.104 ug/ml
7) N-EICOSANE	12.607	18641472	50.325 ug/ml
8) N-DOCOSANE	13.804	18292307	49.317 ug/ml
10) N-TETRACOSANE	14.907	18037430	48.177 ug/ml
11) N-HEXACOSANE	15.926	17390835	46.860 ug/ml
12) N-OCTACOSANE	16.876	17018714	45.686 ug/ml
13) N-TRIACONTANE	17.761	16682780	44.733 ug/ml
14) N-DOTRIACONTANE	18.591	15394685	43.687 ug/ml
15) N-TETRATRIACONTANE	19.372	12489493	42.283 ug/ml
16) N-HEXATRIACONTANE	20.108	9242342	40.996 ug/ml
17) N-OCTATRIACONTANE	20.850	7302022	40.092 ug/ml
18) N-TETRACONTANE	21.778	6526902	39.164 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE083120\
Data File : FE032656.D
Signal(s) : FID1B.CH
Acq On : 31 Aug 2020 14:36
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Sep 01 00:37:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
Quant Title :
QLast Update : Thu Aug 13 11:46:52 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

