

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE083123\
 Data File : FE044072.D
 Signal(s) : FID1B.ch
 Acq On : 31 Aug 2023 21:00
 Operator : YP\AJ
 Sample : 03572-07
 Misc : 1 PPM WATER LOD
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2023

Integration File: autoint1.e
 Quant Time: Sep 01 00:27:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE083023.M
 Quant Title :
 QLast Update : Thu Aug 31 02:43:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.023	2400255	16.372 ug/ml
Target Compounds			
2) N-DECANE	4.571	148633	0.882 ug/ml
3) N-DODECANE	6.735	150018	0.881 ug/ml
4) N-TETRADECANE	8.562	158519	0.922 ug/ml
5) N-HEXADECANE	10.169	152028	0.882 ug/ml
6) N-OCTADECANE	11.613	154599	0.890 ug/ml
7) N-EICOSANE	12.924	153896	0.901 ug/ml
8) N-DOCOSANE	14.122	149912	0.892 ug/ml
10) N-TETRACOSANE	15.226	150053	0.909 ug/ml
11) N-HEXACOSANE	16.248	147451	0.917 ug/ml
12) N-OCTACOSANE	17.198	159017	1.017 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE083123\
Data File : FE044072.D
Signal(s) : FID1B.ch
Acq On : 31 Aug 2023 21:00
Operator : YP\AJ
Sample : 03572-07
Misc : 1 PPM WATER LOD
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
LOD-MDL-WATER-01-QT3-2023

Integration File: autoint1.e
Quant Time: Sep 01 00:27:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE083023.M
Quant Title :
QLast Update : Thu Aug 31 02:43:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

