

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE090120\
 Data File : FE032670.D
 Signal(s) : FID1B.CH
 Acq On : 01 Sep 2020 17:45
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 02 01:07:30 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
 Quant Title :
 QLast Update : Thu Aug 13 11:46:52 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.699	16019826	49.778 ug/ml
Target Compounds			
2) N-DECANE	4.447	19447758	54.162 ug/ml
3) N-DODECANE	6.462	19533264	54.038 ug/ml
4) N-TETRADECANE	8.255	19387713	53.682 ug/ml
5) N-HEXADECANE	9.856	19275451	52.839 ug/ml
6) N-OCTADECANE	11.298	19378555	52.167 ug/ml
7) N-EICOSANE	12.606	19043969	51.412 ug/ml
8) N-DOCOSANE	13.803	18728340	50.493 ug/ml
10) N-TETRACOSANE	14.906	18477963	49.354 ug/ml
11) N-HEXACOSANE	15.925	17845170	48.084 ug/ml
12) N-OCTACOSANE	16.873	17512172	47.010 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE090120\
Data File : FE032670.D
Signal(s) : FID1B.CH
Acq On : 01 Sep 2020 17:45
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 02 01:07:30 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE081120.M
Quant Title :
QLast Update : Thu Aug 13 11:46:52 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

