

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE090619\
 Data File : FE030331.D
 Signal(s) : FID1B.CH
 Acq On : 06 Sep 2019 12:29
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 07 00:23:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE080919.M
 Quant Title :
 QLast Update : Fri Aug 09 14:35:21 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.574	10003068	51.343 ug/ml
Target Compounds			
1) N-OCTANE	2.344	76662562	431.098 ug/ml
2) N-NONANE	3.313	79646269	442.687 ug/ml
3) N-DECANE	4.370	122990191	442.848 ug/ml
4) N-UNDECANE	5.399	123796447	661.638 ug/ml
5) N-DODECANE	6.380	123087413	658.492 ug/ml
6) N-TETRADECANE	8.178	123749751	652.318 ug/ml
7) N-HEXADECANE	9.774	102460458	538.730 ug/ml
8) N-OCTADECANE	11.195	50571937	266.998 ug/ml
9) N-EICOSANE	12.485	19959124	105.704 ug/ml
10) N-TETRACOSANE	14.789	19252551	102.539 ug/ml
11) N-OCTACOSANE	16.751	9207866	49.760 ug/ml
12) N-DOTRIACONTANE	18.471	8799173	48.183 ug/ml
14) N-HEXATRIACONTANE	19.996	8248100	49.067 ug/ml
15) N-TETRACONTANE	21.578	7446239	47.779 ug/ml
16) N-TETRTETRACOOCTANE	24.147	4992502	33.712 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE090619\
Data File : FE030331.D
Signal(s) : FID1B.CH
Acq On : 06 Sep 2019 12:29
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 07 00:23:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE080919.M
Quant Title :
QLast Update : Fri Aug 09 14:35:21 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 uL
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

