

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE091024\  
 Data File : FE049988.D  
 Signal(s) : FID1B.ch  
 Acq On : 09 Sep 2024 23:21  
 Operator : YP\AJ  
 Sample : 10 TRPH STD  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 10 TRPH STD

Integration File: autoint1.e  
 Quant Time: Sep 10 08:22:16 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\FE091024.M  
 Quant Title :  
 QLast Update : Tue Sep 10 08:22:07 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1uL  
 Signal Phase : Rxi-1ms  
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units
-----			
System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.960	1150485	9.875 ug/ml
Target Compounds			
1) N-OCTANE	2.001	1306973	10.068 ug/ml
2) N-DECANE	4.514	1275899	9.860 ug/ml
3) N-DODECANE	6.684	1270585	9.744 ug/ml
4) N-TETRADECANE	8.511	1301245	9.817 ug/ml
5) N-HEXADECANE	10.117	1299015	9.794 ug/ml
6) N-OCTADECANE	11.558	1322831	9.803 ug/ml
7) N-EICOSANE	12.866	1337873	9.845 ug/ml
8) N-DOCOSANE	14.063	1297914	9.864 ug/ml
10) N-TETRACOSANE	15.164	1301885	9.922 ug/ml
11) N-HEXACOSANE	16.183	1294828	9.880 ug/ml
12) N-OCTACOSANE	17.132	1303166	9.902 ug/ml
13) N-TRIACONTANE	18.018	1366183	9.999 ug/ml
14) N-DOTRIACONTANE	18.849	1410343	10.085 ug/ml
15) N-TETRATRIACONTANE	19.632	1445570	10.159 ug/ml
16) N-HEXATRIACONTANE	20.372	1413458	10.208 ug/ml
17) N-OCTATRIACONTANE	21.098	1425440	10.434 ug/ml
18) N-TETRACONTANE	21.989	1368152	10.533 ug/ml
-----			

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE091024\  
 Data File : FE049988.D  
 Signal(s) : FID1B.ch  
 Acq On : 09 Sep 2024 23:21  
 Operator : YP\AJ  
 Sample : 10 TRPH STD  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 10 TRPH STD

Integration File: autoint1.e  
 Quant Time: Sep 10 08:22:16 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\FE091024.M  
 Quant Title :  
 QLast Update : Tue Sep 10 08:22:07 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1uL  
 Signal Phase : Rxi-1ms  
 Signal Info : 20mx0.18mmx0.18um

