

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091318\
Data File : FE028636.D
Signal(s) : FID1B.CH
Acq On : 13 Sep 2018 12:23
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 13 13:54:19 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE082118.M
Quant Title :
QLast Update : Tue Aug 21 14:35:20 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
Target Compounds				
5)	N-DODECANE	6.683	4997346	33.163 ug/ml
11)	N-OCTACOSANE	17.152	4196355	18.878 ug/ml
12)	N-DOTRIACONTANE	18.872	378405	2.064 ug/ml
14)	N-HEXATRIACONTANE	20.400	195262	1.482 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091318\
Data File : FE028636.D
Signal(s) : FID1B.CH
Acq On : 13 Sep 2018 12:23
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 13 13:54:19 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE082118.M
Quant Title :
QLast Update : Tue Aug 21 14:35:20 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um






#5 N-DODECANE

R.T.: 6.683 min
Delta R.T.: -0.005 min
Response: 4997346
Conc: 33.16 ug/ml




#11 N-OCTACOSANE

R.T.: 17.152 min
Delta R.T.: 0.004 min
Response: 4196355
Conc: 18.88 ug/ml




#12 N-DOTRIACONTANE

R.T.: 18.872 min
Delta R.T.: -0.002 min
Response: 378405
Conc: 2.06 ug/ml



#13 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T.: 14.966 min
Response: 0
Conc: N.D.



#14 N-HEXATRIACONTANE

R.T.: 20.400 min
Delta R.T.: -0.002 min
Response: 195262
Conc: 1.48 ug/ml