

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091624AL\
 Data File : FE050131.D
 Signal(s) : FID1B.ch
 Acq On : 16 Sep 2024 20:54
 Operator : YP\AJ
 Sample : P3979-02DL 10X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :

FID_E

ClientSampleId :

EO-03-091224-E2DL

Manual Integrations**APPROVED**

Reviewed By :Yogesh Patel 09/17/2024

Supervised By :Ankita Jodhani 09/17/2024

Integration File: autoint1.e

Quant Time: Sep 17 01:40:55 2024

Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 091024.M

Quant Title : GC Extractables

QLast Update : Mon Sep 09 17:11:47 2024

Response via : Initial Calibration

Integrator: ChemStation

Volume Inj. : 1 ul

Signal Phase : Rxi-1ms

Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.875	466812	3.138 ug/mlm
Spiked Amount 50.000		Recovery =	6.28%
12) S 1-chlorooctadecane (S...	13.310	446615	3.976 ug/ml
Spiked Amount 50.000		Recovery =	7.95%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091624AL\
Data File : FE050131.D
Signal(s) : FID1B.ch
Acq On : 16 Sep 2024 20:54
Operator : YP\AJ
Sample : P3979-02DL 10X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :

FID_E

ClientSampleId :

EO-03-091224-E2DL

Manual Integrations


APPROVED


Reviewed By :Yogesh Patel 09/17/2024

Supervised By :Ankita Jodhani 09/17/2024

Integration File: autoint1.e
Quant Time: Sep 17 01:40:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 091024.
Quant Title : GC Extractables
QLast Update : Mon Sep 09 17:11:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.875 min
Delta R.T.: -0.002 min
Response: 466812
Conc: 3.14 ug/ml

Instrument :
FID_E
ClientSampleId :
EO-03-091224-E2DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/17/2024
Supervised By :Ankita Jodhani 09/17/2024



#12 1-chlorooctadecane (SURR)

R.T.: 13.310 min
Delta R.T.: -0.002 min
Response: 446615
Conc: 3.98 ug/ml