

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091720\
 Data File : FE032747.D
 Signal(s) : FID1B.CH
 Acq On : 17 Sep 2020 16:54
 Operator : DD\AJ
 Sample : L4020-01MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 ETGI-362MS

Integration File: autoint1.e
 Quant Time: Sep 18 00:37:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
 Quant Title :
 QLast Update : Tue Sep 15 03:21:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.687	7798512	24.891 ug/ml
Target Compounds			
2) N-DECANE	4.429	6127668	18.218 ug/ml
3) N-DODECANE	6.441	6269391	18.530 ug/ml
4) N-TETRADECANE	8.236	6509059	19.318 ug/ml
5) N-HEXADECANE	9.835	6577711	18.998 ug/ml
6) N-OCTADECANE	11.278	6644277	18.936 ug/ml
7) N-EICOSANE	12.590	7278737	20.816 ug/ml
8) N-DOCOSANE	13.786	7349867	21.115 ug/ml
10) N-TETRACOSANE	14.894	16459216	46.828 ug/ml
11) N-HEXACOSANE	15.914	7192624	20.627 ug/ml
12) N-OCTACOSANE	16.862	6766024	19.503 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091720\
Data File : FE032747.D
Signal(s) : FID1B.CH
Acq On : 17 Sep 2020 16:54
Operator : DD\AJ
Sample : L4020-01MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
ETGI-362MS

Integration File: autoint1.e
Quant Time: Sep 18 00:37:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
Quant Title :
QLast Update : Tue Sep 15 03:21:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

