

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091720\
 Data File : FE032755.D
 Signal(s) : FID1B.CH
 Acq On : 17 Sep 2020 21:24
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Sep 18 00:39:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
 Quant Title :
 QLast Update : Tue Sep 15 03:21:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.686	14830373	47.336 ug/ml
Target Compounds			
2) N-DECANE	4.433	16446356	48.897 ug/ml
3) N-DODECANE	6.447	16607567	49.085 ug/ml
4) N-TETRADECANE	8.240	16535659	49.076 ug/ml
5) N-HEXADECANE	9.842	16857698	48.688 ug/ml
6) N-OCTADECANE	11.282	16971048	48.366 ug/ml
7) N-EICOSANE	12.592	16789980	48.016 ug/ml
8) N-DOCOSANE	13.790	16569664	47.601 ug/ml
10) N-TETRACOSANE	14.893	16521478	47.005 ug/ml
11) N-HEXACOSANE	15.912	16154137	46.326 ug/ml
12) N-OCTACOSANE	16.862	15843321	45.669 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE091720\
Data File : FE032755.D
Signal(s) : FID1B.CH
Acq On : 17 Sep 2020 21:24
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Sep 18 00:39:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
Quant Title :
QLast Update : Tue Sep 15 03:21:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

