

nteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE09182
Data File : FE050181.D
Signal(s) : FID1B.ch
Acq On : 18 Sep 2024 19:11
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 091024.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.262	3.241	3.339	BBA	261429	2562177	86.08%	4.400%
2	4.513	4.448	4.614	BB	267411	2702483	90.79%	4.641%
3	6.222	6.161	6.398	BB	254943	2766043	92.93%	4.750%
4	6.680	6.651	6.794	BB	270615	2732098	91.79%	4.692%
5	7.318	7.288	7.498	BV	235094	2626451	88.24%	4.511%
6	8.507	8.471	8.579	BBA	254665	2704306	90.85%	4.644%
7	10.113	10.074	10.199	BBA	251894	2790386	93.75%	4.792%
8	11.555	11.511	11.639	BBA	247502	2809869	94.40%	4.826%
9	11.874	11.798	11.978	BB	271302	2976517	100.00%	5.112%
10	12.863	12.831	12.944	BB	223798	2694695	90.53%	4.628%
11	13.309	13.224	13.401	BV	193112	2271813	76.32%	3.902%
12	13.475	13.438	13.571	VB	217459	2643569	88.81%	4.540%
13	14.060	14.025	14.158	VB	216199	2623982	88.16%	4.506%
14	15.162	15.071	15.254	BB	200487	2626807	88.25%	4.511%
15	16.182	16.141	16.249	BBA	197911	2591229	87.06%	4.450%
16	17.130	17.038	17.198	BB	194448	2651839	89.09%	4.554%
17	18.017	17.931	18.071	PV	216259	2878992	96.72%	4.944%
18	18.848	18.801	18.911	BV	209963	2907220	97.67%	4.993%
19	19.631	19.581	19.704	BV	191643	2678856	90.00%	4.601%
20	20.371	20.081	20.431	PBA	169585	2737282	91.96%	4.701%
21	21.097	21.051	21.169	BBA	137916	2282060	76.67%	3.919%
22	21.990	21.931	22.046	BV	87338	1969962	66.18%	3.383%
Sum of corrected areas:							58228637	

Aliphatic EPH 091024.M Thu Sep 19 01:59:35 2024