

rters

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE09182
Data File : FE050197.D
Signal(s) : FID1B.ch
Acq On : 19 Sep 2024 04:15
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 091024.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.260	3.241	3.339	BBA	266946	2569588	85.56%	4.371%
2	4.511	4.448	4.618	BB	274920	2758905	91.86%	4.693%
3	6.220	6.168	6.418	BB	252734	2821335	93.94%	4.799%
4	6.678	6.651	6.781	BB	269520	2772432	92.31%	4.716%
5	7.316	7.284	7.498	PB	231520	2658085	88.51%	4.521%
6	8.506	8.471	8.579	BBA	262873	2745920	91.43%	4.671%
7	10.112	10.071	10.199	BBA	257564	2823176	94.00%	4.802%
8	11.554	11.518	11.639	BBA	253138	2844308	94.71%	4.838%
9	11.873	11.794	11.974	BB	267186	3003259	100.00%	5.109%
10	12.861	12.831	12.938	BB	233395	2716425	90.45%	4.621%
11	13.308	13.271	13.394	BB	190602	2280132	75.92%	3.879%
12	13.472	13.434	13.564	BB	223326	2665130	88.74%	4.533%
13	14.058	14.023	14.148	VB	219264	2646678	88.13%	4.502%
14	15.160	15.081	15.254	BB	203622	2652815	88.33%	4.512%
15	16.179	16.141	16.249	BBA	199438	2617867	87.17%	4.453%
16	17.128	17.034	17.248	BB	193719	2705361	90.08%	4.602%
17	18.015	17.928	18.071	PB	206121	2889399	96.21%	4.915%
18	18.846	18.801	18.914	BV	221507	2967248	98.80%	5.047%
19	19.628	19.581	19.718	BV	183874	2734292	91.04%	4.651%
20	20.368	20.301	20.431	VBA	163900	2470292	82.25%	4.202%
21	21.093	21.051	21.169	BBA	136942	2317223	77.16%	3.942%
22	21.983	21.931	22.074	BV	92132	2128774	70.88%	3.621%
Sum of corrected areas:						58788643		

Aliphatic EPH 091024.M Thu Sep 19 04:56:45 2024