

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092222AL\
 Data File : FE037743.D
 Signal(s) : FID1B.ch
 Acq On : 22 Sep 2022 23:40
 Operator : YP\AJ
 Sample : N4753-01DL 10X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 EO-01-092122DL

Integration File: autoint1.e
 Quant Time: Sep 23 03:45:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 082922.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 30 06:54:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.626	384963	2.278 ug/ml
Spiked Amount	50.000	Recovery =	4.56%
12) S 1-chlorooctadecane (S...	13.049	300571	2.335 ug/ml
Spiked Amount	50.000	Recovery =	4.67%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092222AL\
Data File : FE037743.D
Signal(s) : FID1B.ch
Acq On : 22 Sep 2022 23:40
Operator : YP\AJ
Sample : N4753-01DL 10X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
EO-01-092122DL

Integration File: autoint1.e
Quant Time: Sep 23 03:45:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 082922.M
Quant Title : GC Extractables
QLast Update : Tue Aug 30 06:54:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

