



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

G Environmental

Project Name: 61 Laurel Street

Project # N/A

Order ID # Q3108

Test Name: SVOCMS Group1

A. Number of Samples and Date of Receipt:

1 Water sample was received on 09/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:

. This data package contains results for SVOCMS Group1(8270E).

C. Analytical Techniques:

SVOCMS Group1 : The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike for {PB169719BS} with File ID: BP025774.D met requirements for all compounds except for Acenaphthylene[78%], Hexachloroethane[74%], are marginally biased low, therefore no corrective action was taken.

The Blank Spike Duplicate met requirements for all compounds except following

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:

SVOCMS Group1 : The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

F. Manual Integration Comments:



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____