

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE092524AL\  
 Data File : FE050360.D  
 Signal(s) : FID1B.ch  
 Acq On : 25 Sep 2024 21:30  
 Operator : YP\AJ  
 Sample : P4168-03DL 5X  
 Misc :  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :  
 SU-04-09242024DL

Integration File: autoint1.e  
 Quant Time: Sep 26 01:18:58 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 091024.M  
 Quant Title : GC Extractables  
 QLast Update : Mon Sep 09 17:11:47 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units  |
|--------------------------------|--------|------------|-------------|
| -----                          |        |            |             |
| System Monitoring Compounds    |        |            |             |
| 9) S ortho-Terphenyl (SURR)    | 11.867 | 845625     | 5.684 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 11.37%      |
| 12) S 1-chlorooctadecane (S... | 13.301 | 633297     | 5.638 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 11.28%      |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE092524AL\  
Data File : FE050360.D  
Signal(s) : FID1B.ch  
Acq On : 25 Sep 2024 21:30  
Operator : YP\AJ  
Sample : P4168-03DL 5X  
Misc :  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
SU-04-09242024DL

Integration File: autoint1.e  
Quant Time: Sep 26 01:18:58 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 091024.M  
Quant Title : GC Extractables  
QLast Update : Mon Sep 09 17:11:47 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.867 min  
Delta R.T.: -0.010 min  
Response: 845625  
Conc: 5.68 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
SU-04-09242024DL



#12 1-chlorooctadecane (SURR)

R.T.: 13.301 min  
Delta R.T.: -0.010 min  
Response: 633297  
Conc: 5.64 ug/ml