

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092618\
 Data File : FE028807.D
 Signal(s) : FID1B.CH
 Acq On : 27 Sep 2018 00:44
 Operator : AJ\SJ
 Sample : J5074-15MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 27 04:16:59 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
 Quant Title :
 QLast Update : Fri Sep 21 07:24:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.942	3779684	18.300 ug/ml
Target Compounds			
1) N-OCTANE	2.528	1542075	15.536 ug/ml
3) N-DECANE	4.631	2859060	17.865 ug/ml
5) N-DODECANE	6.659	3488030	28.750 ug/ml
6) N-TETRADECANE	8.465	4145585	31.255 ug/ml
7) N-HEXADECANE	10.072	4227998	28.247 ug/ml
8) N-OCTADECANE	11.520	4349763	26.481 ug/ml
9) N-EICOSANE	12.836	4328080	24.678 ug/ml
10) N-TETRACOSANE	15.147	3968792	22.755 ug/ml
11) N-OCTACOSANE	17.126	3783836	24.640 ug/mlm
12) N-DOTRIACONTANE	18.852	4064900	31.290 ug/ml
14) N-HEXATRIACONTANE	20.379	3085139	30.532 ug/ml
15) N-TETRACONTANE	22.184	2952996	40.840 ug/ml
16) N-TETRTETRACOOCTANE	25.266	159966	3.946 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092618\
Data File : FE028807.D
Signal(s) : FID1B.CH
Acq On : 27 Sep 2018 00:44
Operator : AJ\SJ
Sample : J5074-15MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 27 04:16:59 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
Quant Title :
QLast Update : Fri Sep 21 07:24:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

