

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092618\
 Data File : FE028813.D
 Signal(s) : FID1B.CH
 Acq On : 27 Sep 2018 4:23
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 27 08:39:28 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
 Quant Title :
 QLast Update : Fri Sep 21 07:24:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.950	11057342	53.537 ug/ml
Target Compounds			
1) N-OCTANE	2.549	39323090	396.180 ug/ml
2) N-NONANE	3.580	42416365	411.073 ug/ml
3) N-DECANE	4.666	68433240	427.606 ug/ml
4) N-UNDECANE	5.711	76640082	656.440 ug/ml
5) N-DODECANE	6.701	79838306	658.058 ug/ml
6) N-TETRADECANE	8.510	85257202	642.775 ug/ml
7) N-HEXADECANE	10.116	74237721	495.982 ug/ml
8) N-OCTADECANE	11.548	36613024	222.899 ug/ml
9) N-EICOSANE	12.847	13929403	79.422 ug/ml
10) N-TETRACOSANE	15.158	12150108	69.662 ug/ml
11) N-OCTACOSANE	17.128	6052049	39.411 ug/ml
12) N-DOTRIACONTANE	18.853	5874076	45.216 ug/ml
14) N-HEXATRIACONTANE	20.380	4729955	46.809 ug/ml
15) N-TETRACONTANE	22.183	3551280	49.115 ug/ml
16) N-TETRATETRAOCTANE	25.274	2594304	63.991 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092618\
Data File : FE028813.D
Signal(s) : FID1B.CH
Acq On : 27 Sep 2018 4:23
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 27 08:39:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
Quant Title :
QLast Update : Fri Sep 21 07:24:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

