

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092619\
 Data File : FE030569.D
 Signal(s) : FID1B.CH
 Acq On : 26 Sep 2019 12:19
 Operator : DD\AJ
 Sample : 880 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: sample.e
 Quant Time: Sep 26 12:51:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092619.M
 Quant Title :
 QLast Update : Thu Sep 26 12:02:23 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.676	2925861	9.329 ug/ml
Target Compounds			
1) N-OCTANE	2.380	24298487	75.120 ug/ml
2) N-NONANE	3.370	24567187	75.248 ug/ml
3) N-DECANE	4.428	37218089	75.269 ug/ml
4) N-UNDECANE	5.462	37253272	113.110 ug/ml
5) N-DODECANE	6.444	36876003	113.149 ug/ml
6) N-TETRADECANE	8.240	36630695	113.046 ug/ml
7) N-HEXADECANE	9.840	30056074	93.960 ug/ml
8) N-OCTADECANE	11.274	14775833	46.892 ug/ml
9) N-EICOSANE	12.579	5834989	18.719 ug/ml
10) N-TETRACOSANE	14.676	2925861	18.658 ug/ml
11) N-OCTACOSANE	16.851	2778915	9.435 ug/ml
12) N-DOTRIACONTANE	18.572	2672240	9.480 ug/ml
14) N-HEXATRIACONTANE	20.096	2431092	9.286 ug/ml
15) N-TETRACONTANE	21.734	2165256	10.311 ug/ml
16) N-TETRTETRACOOCTANE	24.496	1787329	11.177 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE092619\
Data File : FE030569.D
Signal(s) : FID1B.CH
Acq On : 26 Sep 2019 12:19
Operator : DD\AJ
Sample : 880 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: sample.e
Quant Time: Sep 26 12:51:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092619.M
Quant Title :
QLast Update : Thu Sep 26 12:02:23 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

