

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100118\
 Data File : FE028853.D
 Signal(s) : FID1B.CH
 Acq On : 01 Oct 2018 22:49
 Operator : AJ\SJ
 Sample : J5170-28
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 02 05:59:48 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
 Quant Title :
 QLast Update : Fri Sep 21 07:24:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.932	3338055	16.162 ug/ml
Target Compounds			
6) N-TETRADECANE	8.524	776190	5.852 ug/ml
11) N-OCTACOSANE	17.111	173079	1.127 ug/ml
12) N-DOTRIACONTANE	18.836	303886	2.339 ug/ml
14) N-HEXATRIACONTANE	20.402	209585	2.074 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100118\
Data File : FE028853.D
Signal(s) : FID1B.CH
Acq On : 01 Oct 2018 22:49
Operator : AJ\SJ
Sample : J5170-28
Misc :
ALS Vial : 20 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 02 05:59:48 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
Quant Title :
QLast Update : Fri Sep 21 07:24:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um






#6 N-TETRADECANE

R.T.: 8.524 min
Delta R.T.: 0.049 min
Response: 776190
Conc: 5.85 ug/ml




#11 N-OCTACOSANE

R.T.: 17.111 min
Delta R.T.: -0.016 min
Response: 173079
Conc: 1.13 ug/ml




#12 N-DOTRIACONTANE

R.T.: 18.836 min
Delta R.T.: -0.016 min
Response: 303886
Conc: 2.34 ug/ml



#13 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.932 min
Delta R.T.: -0.012 min
Response: 3338055
Conc: 16.16 ug/ml



#14 N-HEXATRIACONTANE

R.T.: 20.402 min
Delta R.T.: 0.021 min
Response: 209585
Conc: 2.07 ug/ml