

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100119\
 Data File : FE030628.D
 Signal(s) : FID1B.CH
 Acq On : 01 Oct 2019 17:54
 Operator : DD\AJ
 Sample : K5094-04MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 MC-17-9.0-9.5MS

Integration File: autoint1.e
 Quant Time: Oct 02 00:39:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092419.M
 Quant Title :
 QLast Update : Wed Sep 25 02:50:34 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.680	5586479	18.020 ug/ml
Target Compounds			
2) N-DECANE	4.414	7844316	20.551 ug/ml
3) N-DODECANE	6.429	8011848	20.933 ug/ml
4) N-TETRADECANE	8.226	8076103	21.479 ug/ml
5) N-HEXADECANE	9.828	8160392	21.858 ug/ml
6) N-OCTADECANE	11.271	8061460	21.997 ug/ml
7) N-EICOSANE	12.582	7917654	21.701 ug/ml
8) N-DOCOSANE	13.782	7811678	21.765 ug/ml
10) N-TETRACOSANE	14.885	7771092	21.857 ug/ml
11) N-HEXACOSANE	15.907	7634587	21.788 ug/ml
12) N-OCTACOSANE	16.858	7682311	22.324 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100119\
Data File : FE030628.D
Signal(s) : FID1B.CH
Acq On : 01 Oct 2019 17:54
Operator : DD\AJ
Sample : K5094-04MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
MC-17-9.0-9.5MS

Integration File: autoint1.e
Quant Time: Oct 02 00:39:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092419.M
Quant Title :
QLast Update : Wed Sep 25 02:50:34 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

