

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100120\
 Data File : FE032856.D
 Signal(s) : FID1B.CH
 Acq On : 01 Oct 2020 19:55
 Operator : DD\AJ
 Sample : PB132019BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 PB132019BS

Integration File: autoint1.e
 Quant Time: Oct 02 01:22:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
 Quant Title :
 QLast Update : Tue Sep 15 03:21:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.677	5592553	17.850 ug/ml
Target Compounds			
2) N-DECANE	4.420	5978183	17.774 ug/ml
3) N-DODECANE	6.433	6113907	18.070 ug/ml
4) N-TETRADECANE	8.228	6284962	18.653 ug/ml
5) N-HEXADECANE	9.828	6298891	18.192 ug/ml
6) N-OCTADECANE	11.270	6356289	18.115 ug/ml
7) N-EICOSANE	12.579	6400716	18.305 ug/ml
8) N-DOCOSANE	13.777	6299047	18.096 ug/ml
10) N-TETRACOSANE	14.881	6340058	18.038 ug/ml
11) N-HEXACOSANE	15.902	5912884	16.957 ug/ml
12) N-OCTACOSANE	16.852	5801677	16.724 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100120\
Data File : FE032856.D
Signal(s) : FID1B.CH
Acq On : 01 Oct 2020 19:55
Operator : DD\AJ
Sample : PB132019BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
PB132019BS

Integration File: autoint1.e
Quant Time: Oct 02 01:22:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
Quant Title :
QLast Update : Tue Sep 15 03:21:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

