

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100218\
 Data File : FE028869.D
 Signal(s) : FID1B.CH
 Acq On : 02 Oct 2018 20:36
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 03 06:00:53 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
 Quant Title :
 QLast Update : Fri Sep 21 07:24:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.940	10451948	50.606 ug/ml
Target Compounds			
1) N-OCTANE	2.546	38676561	389.666 ug/ml
2) N-NONANE	3.575	42271599	409.670 ug/ml
3) N-DECANE	4.661	68252162	426.475 ug/ml
4) N-UNDECANE	5.704	75409369	645.898 ug/ml
5) N-DODECANE	6.694	77750919	640.853 ug/ml
6) N-TETRADECANE	8.501	81183892	612.065 ug/ml
7) N-HEXADECANE	10.107	69136032	461.898 ug/ml
8) N-OCTADECANE	11.538	33290541	202.672 ug/ml
9) N-EICOSANE	12.837	12758821	72.748 ug/ml
10) N-TETRACOSANE	15.148	11688425	67.015 ug/ml
11) N-OCTACOSANE	17.118	5928968	38.609 ug/ml
12) N-DOTRIACONTANE	18.845	5962552	45.897 ug/ml
14) N-HEXATRIACONTANE	20.371	4644000	45.959 ug/ml
15) N-TETRACONTANE	22.166	3411629	47.183 ug/ml
16) N-TETRATETRAOCTANE	25.243	2432593	60.002 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100218\
Data File : FE028869.D
Signal(s) : FID1B.CH
Acq On : 02 Oct 2018 20:36
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 03 06:00:53 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
Quant Title :
QLast Update : Fri Sep 21 07:24:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

