

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100418\
 Data File : FE028872.D
 Signal(s) : FID1B.CH
 Acq On : 04 Oct 2018 10:30
 Operator : DD\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 15:01:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092218.M
 Quant Title :
 QLast Update : Fri Sep 21 08:15:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.940	9765890	45.488 ug/ml
Target Compounds			
2) N-DECANE	4.631	10386635	73.514 ug/ml
3) N-DODECANE	6.660	11759738	74.255 ug/ml
4) N-TETRADECANE	8.465	12722239	68.738 ug/ml
5) N-HEXADECANE	10.074	13103262	61.631 ug/ml
6) N-OCTADECANE	11.522	12882455	54.218 ug/ml
7) N-EICOSANE	12.837	12143901	49.096 ug/ml
8) N-DOCOSANE	14.040	10525775	42.201 ug/ml
10) N-TETRACOSANE	15.148	10902518	44.522 ug/ml
11) N-HEXACOSANE	16.172	10721031	45.774 ug/ml
12) N-OCTACOSANE	17.126	10820278	49.663 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100418\
Data File : FE028872.D
Signal(s) : FID1B.CH
Acq On : 04 Oct 2018 10:30
Operator : DD\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 15:01:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092218.M
Quant Title :
QLast Update : Fri Sep 21 08:15:58 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

