

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100418\
 Data File : FE028875.D
 Signal(s) : FID1B.CH
 Acq On : 04 Oct 2018 12:10
 Operator : DD\SJ
 Sample : J5164-07
 Misc : FID_E LOD WATER 1 PPM
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 15:01:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092218.M
 Quant Title :
 QLast Update : Fri Sep 21 08:15:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.930	2692430	12.541 ug/ml
Target Compounds			
2) N-DECANE	4.624	147238	1.042 ug/ml
3) N-DODECANE	6.650	124536	0.786 ug/mlm
4) N-TETRADECANE	8.452	166256	0.898 ug/ml
5) N-HEXADECANE	10.060	177151	0.833 ug/ml
6) N-OCTADECANE	11.507	177077	0.745 ug/ml
7) N-EICOSANE	12.822	181117	0.732 ug/ml
8) N-DOCOSANE	14.024	183703	0.737 ug/mlm
10) N-TETRACOSANE	15.133	169247	0.691 ug/ml
11) N-HEXACOSANE	16.157	173276	0.740 ug/ml
12) N-OCTACOSANE	17.110	208228	0.956 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100418\
Data File : FE028875.D
Signal(s) : FID1B.CH
Acq On : 04 Oct 2018 12:10
Operator : DD\SJ
Sample : J5164-07
Misc : FID_E LOD WATER 1 PPM
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 15:01:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092218.M
Quant Title :
QLast Update : Fri Sep 21 08:15:58 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

