

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100418\
 Data File : FE028877.D
 Signal(s) : FID1B.CH
 Acq On : 04 Oct 2018 13:16
 Operator : DD\SJ
 Sample : J5164-09
 Misc : FID_E MDL WATER 1 PPM
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 04 15:01:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092218.M
 Quant Title :
 QLast Update : Fri Sep 21 08:15:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.931	3057186	14.240 ug/ml
Target Compounds				
2)	N-DECANE	4.625	191823	1.358 ug/ml
3)	N-DODECANE	6.649	177359	1.120 ug/ml
4)	N-TETRADECANE	8.452	195620	1.057 ug/ml
5)	N-HEXADECANE	10.059	200001	0.941 ug/ml
6)	N-OCTADECANE	11.507	199420	0.839 ug/ml
7)	N-EICOSANE	12.822	200458	0.810 ug/ml
8)	N-DOCOSANE	14.024	210024	0.842 ug/mlm
10)	N-TETRACOSANE	15.133	200030	0.817 ug/ml
11)	N-HEXACOSANE	16.156	212425	0.907 ug/ml
12)	N-OCTACOSANE	17.110	259083	1.189 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100418\
Data File : FE028877.D
Signal(s) : FID1B.CH
Acq On : 04 Oct 2018 13:16
Operator : DD\SJ
Sample : J5164-09
Misc : FID_E MDL WATER 1 PPM
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 04 15:01:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE092218.M
Quant Title :
QLast Update : Fri Sep 21 08:15:58 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

