

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100522\
 Data File : FE037968.D
 Signal(s) : FID1B.ch
 Acq On : 05 Oct 2022 23:11
 Operator : YP\AJ
 Sample : N4955-08
 Misc : 5 PPM WATER LOQ
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 LOQ-WATER-02-QT4-2022

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/06/2022
 Supervised By :Ankita Jodhani 10/06/2022

Integration File: autoint1.e
 Quant Time: Oct 06 05:48:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100522.M
 Quant Title :
 QLast Update : Wed Oct 05 16:55:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.668	2241391	17.256 ug/ml
Target Compounds			
2) N-DECANE	4.260	499518	3.733 ug/mlm
3) N-DODECANE	6.433	519282	3.714 ug/mlm
4) N-TETRADECANE	8.251	574834	3.956 ug/mlm
5) N-HEXADECANE	9.849	596409	3.977 ug/ml
6) N-OCTADECANE	11.282	620598	4.088 ug/ml
7) N-EICOSANE	12.582	634226	4.166 ug/ml
8) N-DOCOSANE	13.773	641836	4.309 ug/ml
10) N-TETRACOSANE	14.869	653628	4.396 ug/ml
11) N-HEXACOSANE	15.884	631537	4.371 ug/ml
12) N-OCTACOSANE	16.828	615198	4.372 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100522\
Data File : FE037968.D
Signal(s) : FID1B.ch
Acq On : 05 Oct 2022 23:11
Operator : YP\AJ
Sample : N4955-08
Misc : 5 PPM WATER LOQ
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
LOQ-WATER-02-QT4-2022

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/06/2022
Supervised By :Ankita Jodhani 10/06/2022

Integration File: autoint1.e
Quant Time: Oct 06 05:48:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100522.M
Quant Title :
QLast Update : Wed Oct 05 16:55:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

