

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100623\
 Data File : FE044681.D
 Signal(s) : FID1B.ch
 Acq On : 07 Oct 2023 08:43
 Operator : YP\AJ
 Sample : 04699-10
 Misc : 1PPM WATER LOD
 ALS Vial : 48 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 09 00:18:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100323.M
 Quant Title :
 QLast Update : Tue Oct 03 12:58:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.029	3394071	23.134 ug/ml
Target Compounds			
1) N-OCTANE	2.025	240376	1.482 ug/ml
2) N-DECANE	4.552	265836	1.624 ug/ml
3) N-DODECANE	6.721	259623	1.567 ug/ml
4) N-TETRADECANE	8.550	267992	1.613 ug/ml
5) N-HEXADECANE	10.162	273930	1.626 ug/ml
6) N-OCTADECANE	11.609	276078	1.651 ug/ml
7) N-EICOSANE	12.923	270445	1.650 ug/ml
8) N-DOCOSANE	14.125	271760	1.727 ug/ml
10) N-TETRACOSANE	15.232	256822	1.659 ug/ml
11) N-HEXACOSANE	16.256	252628	1.676 ug/ml
12) N-OCTACOSANE	17.208	254974	1.716 ug/ml
13) N-TRIACONTANE	18.097	272241	1.835 ug/ml
14) N-DOTRIACONTANE	18.933	292330	2.017 ug/ml
15) N-TETRATRIACONTANE	19.719	278942	2.181 ug/ml
16) N-HEXATRIACONTANE	20.463	263521	2.495 ug/ml
17) N-OCTATRIACONTANE	21.210	252463	2.537 ug/ml
18) N-TETRACONTANE	22.140	262903	2.574 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100623\
Data File : FE044681.D
Signal(s) : FID1B.ch
Acq On : 07 Oct 2023 08:43
Operator : YP\AJ
Sample : 04699-10
Misc : 1PPM WATER LOD
ALS Vial : 48 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 00:18:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100323.M
Quant Title :
QLast Update : Tue Oct 03 12:58:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

