

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100623\
 Data File : FE044687.D
 Signal(s) : FID1B.ch
 Acq On : 07 Oct 2023 12:16
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 09 00:19:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100323.M
 Quant Title :
 QLast Update : Tue Oct 03 12:58:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.021	6884173	46.923 ug/ml
Target Compounds			
2) N-DECANE	4.552	7744484	47.309 ug/ml
3) N-DODECANE	6.720	7924770	47.824 ug/ml
4) N-TETRADECANE	8.548	7918254	47.667 ug/ml
5) N-HEXADECANE	10.160	7983987	47.398 ug/ml
6) N-OCTADECANE	11.605	7887808	47.181 ug/ml
7) N-EICOSANE	12.918	7719875	47.089 ug/ml
8) N-DOCOSANE	14.120	7402312	47.053 ug/ml
10) N-TETRACOSANE	15.225	7326588	47.315 ug/ml
11) N-HEXACOSANE	16.248	7138289	47.347 ug/ml
12) N-OCTACOSANE	17.200	7016981	47.238 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100623\
Data File : FE044687.D
Signal(s) : FID1B.ch
Acq On : 07 Oct 2023 12:16
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 41 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 00:19:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100323.M
Quant Title :
QLast Update : Tue Oct 03 12:58:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

