

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100623AL\
 Data File : FE044663.D
 Signal(s) : FID1B.ch
 Acq On : 06 Oct 2023 22:01
 Operator : YP\AJ
 Sample : 04778-05
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 07 04:54:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 092523.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 25 11:06:21 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.927	5693686	30.328 ug/ml
Spiked Amount 50.000		Recovery =	60.66%
12) S 1-chlorooctadecane (S...	13.370	4316069	31.726 ug/ml
Spiked Amount 50.000		Recovery =	63.45%
Target Compounds			
10) T n-Eicosane (C20)	12.977f	95340	0.593 ug/ml
14) T n-Tetracosane (C24)	15.201	52402	0.347 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100623AL\
Data File : FE044663.D
Signal(s) : FID1B.ch
Acq On : 06 Oct 2023 22:01
Operator : YP\AJ
Sample : 04778-05
Misc :
ALS Vial : 32 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 07 04:54:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 092523.M
Quant Title : GC Extractables
QLast Update : Mon Sep 25 11:06:21 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.927 min
Delta R.T.: 0.000 min
Response: 5693686
Conc: 30.33 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.977 min
Delta R.T.: 0.057 min
Response: 95340
Conc: 0.59 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.370 min
Delta R.T.: 0.005 min
Response: 4316069
Conc: 31.73 ug/ml



#14 n-Tetracosane (C24)

R.T.: 15.201 min
Delta R.T.: -0.023 min
Response: 52402
Conc: 0.35 ug/ml