

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100823\
Data File : FE044713.D
Signal(s) : FID1B.ch
Acq On : 08 Oct 2023 07:55
Operator : YP\AJ
Sample : PB156132BL
Misc :
ALS Vial : 64 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 00:23:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100323.M
Quant Title :
QLast Update : Tue Oct 03 12:58:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.013	2746138	18.718 ug/ml
Target Compounds			


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100823\
Data File : FE044713.D
Signal(s) : FID1B.ch
Acq On : 08 Oct 2023 07:55
Operator : YP\AJ
Sample : PB156132BL
Misc :
ALS Vial : 64 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 00:23:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100323.M
Quant Title :
QLast Update : Tue Oct 03 12:58:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um





#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 15.013 min
Delta R.T.: 0.000 min
Response: 2746138
Conc: 18.72 ug/ml